










Robby

(D)	Bedienungsanleitung.....	5
(GB USA)	Instructions for Use	10
(F)	Mode d'emploi	15
(I)	Istruzione d'uso	20
(E)	Instrucciones de uso	25





(D) Therapieprodukt

(GB USA) Therapy aid

(F) Produit thérapeutique

(I) Prodotto terapeutico

(E) Producto terapéutico

(P) Produto terapêutico

(NL) Therapieproduct

(DK) Terapiprodukt

(S) Terapiprodukt

(N) Terapiprodukt




- **Bedienungsanleitung beachten!**
- Observe instructions for use!
- Respecter les instructions de service!
- Osservare le istruzioni d'uso!
- Seguir las instrucciones para el uso!
- Leia as instruções de utilização atentamente!
- Let op de gebruiksaanwijzing!
- Betjeningsvejledningen skal overholdes!
- Iakttta bruksanvisningen!
- Følg bruksanvisningen!
- Lue käyttöohje läpi!




- **Nur im Innenbereich verwenden!**
- For indoor use only!
- N'utiliser qu'à l'intérieur !
- Utilizzare solamente all'interno!
- Emplear solamente en el interior!
- O produto é destinado apenas à utilização em interiores!
- Alleen binnen gebruiken!
- Kun til brug inden døre!
- Får endast användas inomhus !
- Må kun brukes innendørs !
- Käytä vain sisätiloissa!



- **Nur unter Aufsicht von Fachpersonal verwenden!**
- Use under supervision of expert staff only!
- N'utiliser que sous la surveillance de personnel spécialisé !
- Utilizzare solamente sotto sorveglianza di personale specializzato!
- Emplear solamente sometido a vigilancia de personal especializado!
- Utilize somente sob vigilância de pessoal qualificado!
- Alleen onder toezicht van vakkundig personeel gebruiken!
- Må kun anvendes under opsyn af fagpersonale!
- Får endast användas under uppsyn av fackpersonall!
- Må kun brukes under oppsyn av fagpersonell!
- Käytä vain hoitohenkilökunnan läsnäollessa!



- **Nicht auf schrägem Untergrund verwenden !**
- Do not use on sloping surface!
- Ne pas utiliser sur sol incliné !
- Non utilizzare su superficie pendente!
- No emplear sobre fondo inclinado!
- Não utilize sobre superfícies inclinadas!
- Niet op een schuine ondergrond gebruiken!
- Må ikke anvendes på skrått underlag!
- Får ej användas på sluttande underlag !
- Må ikke brukes på skrått underlag !
- Älä käytä vinolla pohjalla!



- **Nicht auf unebenem Untergrund verwenden !**
- Do not use on uneven surface!
- Ne pas utiliser sur sol irrégulier !
- Non utilizzare su terreni sconnessi!
- No emplear sobre fondo accidentado!
- Não utilize sobre superfícies acidentadas!
- Niet op een oneffen ondergrond gebruiken!
- Må ikke anvendes på ujevnt underlag!
- Får ej användas på ojämnt underlag !
- Må ikke brukes på ujevnt underlag !
- Älä käytä epätasaisella pohjalla!



Table of Contents	Page
Preface	11
Indications	11
Declaration of Conformity	11
Construction and Function of the Basic Version	11
Setting and Adjustment Options	12
Accessories	12
Headrest	12
Trunk Strap	13
Leg Straps	13
Practical Instructions on Safety, Maintenance and Care	13
Warranty	13
Specifications	13
Spare Parts	14

Preface

Thank you for purchasing the ROBBY bathing and lying support. This brochure describes various setting and adjustment options and offers suggestions for its use.

It will show you how to use the support properly and also explain how to perform simple care and maintenance.

To prevent any damages that could occur due to improper operation and allow for the best possible use, please read the following instructions carefully.

If any questions or problems arise in connection with using this product, please contact your provider.

We hope the ROBBY meets your expectations and helps make your life more comfortable by providing you with trouble-free operation.

Indications

(Source: Catalog and list of therapy aids of German health insurance companies)

Bathing supports are designed for the use in a bath tub. This bathing support can also be used in the shower if it fits.

The ROBBY is intended especially for children from two years of age who lack trunk control or who are inclined to excessive muscular tension (for example, due to cerebral paresis).

Declaration of Conformity

Otto Bock HealthCare GmbH declares with sole responsibility that this product conforms to the requirements of the 93/42 EEC Guidelines.

ROBBY Bathing and Lying Support

The Robby bathing support is a shower and bathing aid that allows you to make various adjustments to best fit the needs of your child.

The Robby bathing support supports the child in both lying and sitting positions.

The hook and loop closure tensioning straps that support the upholstery can be tightened or loosened to better accommodate children with muscular tension.

To keep soap from getting in the child's eyes during hair washing, the headrest can either be turned backwards or the fabric upholstery can be removed from this section to allow for free motion of the head.

The ROBBY has been designed to grow with your child, and has telescoping frame tubes that can easily be modified to change back length and seat depth.

The construction and functions of the Robby bathing support are explained in detail in the following section.

Construction and Function of the Basic Version

The Robby bathing support consists of a lower leg section (1), seat (2), backrest (3) and headrest (4). (Fig. 1)

The frame is covered with two sections of upholstery that overlap each other in the pelvic area and which are connected with hook and loop straps. To accommodate any growth the child may experience, the hook and loop straps can be loosened to allow for greater or lesser upholstery depth.

The angle of the bathing support is set with ratchet-lock joints found on either side (Fig. 2). To unlock and set the ratchet-lock joint, press the push-button (Fig. 3). After setting the position, release the push-button and lock the joint by moving it to the nearest locking position.

The individual sections of the bathing support can be adjusted with the ratchet-lock joints A, B, C and D (Fig. 4).

To adjust the sections, opposing ratchet-lock joint push-buttons must be depressed simultaneously and held during adjustment (Fig. 5).

The easiest way to release the push-button is to press it in with your thumb while using your other fingers to adjust the angle of the support.

Once you've reached the desired position release the push-buttons of the ratcheting-lock joints. The joints will lock automatically.

Upholstery

The tension of the upholstery can be adjusted individually with the hook and loop straps (Fig. 6).

General Handling

If the ROBBY is dry, the easiest way to adjust the support is to butt one section up against your chest while holding it with your hands on the ratcheting-lock joints (Fig. 7). You can facilitate the whole procedure with the upper half of your body.

The setting position shown in Fig. 8 is awkward and requires too much force and is therefore not recommended.

Setting and Adjustment Options

Adjust the bathing support using the procedure described above. Both sitting positions (Fig. 9) and lying positions (Fig. 10). are set in the same way.

Headrest Adjustment

To help prevent soap from getting in the child's eyes release the ratchet-lock joint D and move the headrest backwards (Fig. 11).

Another option is to remove the upholstery from the headrest. In this situation the child's head is completely unsupported and, in case of poor head control, should be supported by hand of the attendant (Fig. 12).

To minimize excessive neck musculature tension, the headrest can also be adjusted forwards in conjunction with an increase in hip flexion,

Setting and Adjustment Options

Length Adjustment of the Back:

To change the length of the back section, loosen the bolts on either side of the back section with a hex cap socket wrench (Fig. 13).

Then loosen the hook and loop straps of both overlapping parts of the upholstery fabric in the pelvic section of the backrest (Fig. 14).

Next pull the headrest upwards and simultaneously hold the seat with the other hand.

When the required position is reached, tighten both bolts again. Make sure that each bolt has been screwed back into the hole it was removed them. (Fig. 15).

Now pull both parts of upholstery fabric tightly again one upon another and connect them with the hook and loop straps (Fig. 16).

Length Adjustment of the Seat:

To change the seat length, loosen the bolts on either side of the seat with a hex cap socket wrench (Fig. 17).

Then loosen the hook and loop straps on both overlapping pelvic sections of upholstery (Fig. 18).

Because the required upholstery in this case comes from the back section, it may be necessary to loosen the lowest tensioning strap.

Pull the foot section so as to set the required length and fix the new setting position with both screws.

Now pull both parts of upholstery fabric again tightly one over the other and connect them with the hook and loop closures. If necessary, connect the loosened tension strap on the backside side again.

Accessories

Figure 21

Headrest

The headrest (Fig. 22) consists of two foam pads with upholstery, which are located on either side of the head (Fig. 23).

Once you've determined the correct headrest height, you can fix it in place with the upholstery length strap (Fig. 24).

To set the correct width of the headrest pads, lift the outside end of the pad to be adjusted and move the pad into the required position (Fig. 25). Subsequently fix the pad by pressing it on the guide belt hook and loop fastener. Repeat this procedure with the other pad.

Trunk Strap

The trunk strap (Fig. 26) offers additional restraint for children who experience involuntary, spontaneous movements, which could otherwise make their position insecure. The trunk belt can be adjusted in height just like the headrest with longitudinal straps attached to the upholstery.

Additional restraint can be provided by reducing the tension of the upholstery to create a larger space for the child.

Another possibility for providing greater trunk support is to use the headrest as a trunk guide. The headrest, with its special shape and size is well suited for this use (Fig. 27).

Leg Straps

The leg straps (Fig. 28) restrain the legs and counteract extension movements, which can make the positioning and care of the child more difficult.

The straps are also height adjustable, just like the headrest and trunk straps (Fig. 29).

If necessary, these straps can be also be used to support and guide the child's thighs.

All accessories, unless installed by the manufacturer, are to be installed by skilled personnel.

Practical Instructions on Safety, Maintenance and Care

- The maximal load capacity is 30 kg for size 1 and 60 kg for size 2.
- Robby is a therapeutic device and can only be used under supervision!
- Use only on a flat, even surface!
- During the installation, or when folding the bathing support, be careful not to pinch your fingers.
- The yoke for back height adjustment, the Ratchet-lock joint B (Fig. 4) and the end of the lower leg rest should always be in contact with the tub.
- Although the position shown in Figure 30 is adjustable, spontaneous movements made by the child (e.g. due to spasticity, cramping, athetoses) may result in a rocking movement of the bathing support, which can cause the child to become unstable or even slip out of the bathing support if restraints aren't used.
- Clean all frame and net parts with mild household detergents only.
- After use the bathing support can be folded to save space (Fig. 19). When the support is folded the headrest should be pointing downwards (Fig. 20). The rubber tube protectors provide protection against slipping and also release any water through the drain holes.
- The mesh material allows for free water flow and can be removed for washing (max. 60° C).
- Prior to washing the device, it is advisable to close the straps and pack the upholstery fabric into a pillow case so that it can be easily unfolded after being washed and prevent from mixing in with other items in the wash.
- If necessary, remove any stray material that accumulates in the hook and loop closures with a nail brush.
- If any damage to the upholstery occurs, replace it immediately.
- No other maintenance is necessary.


Warranty

The Otto Bock Warranty is valid only when the product is used according to the specified conditions and for its intended purposes, following all manufacturer's recommendations.

Technical Data

Measurements and weights	Size 1	Size 2
Overall length of the lying surface	115 cm – 133 cm	147 cm – 173 cm
Seat depth	20 cm – 24 cm	31 cm – 43 cm
Back length	63 cm – 72 cm	78 cm – 94 cm
Width	38 cm	38 cm
Weight	3.8 kg	4.3 kg
Max. load capacity	30 kg	60 kg

Spare Parts



When ordering spare parts, please always indicate the serial number of the product to ensure smooth delivery:

Item	Article number	Description	Quantity
1	HR51041300-027E	Foot frame, size 1, complete	1
	HR51051300-027E	Foot frame, size 2, complete	1
2	HR51041200-027E	Base frame, size 1, complete	1
	HR51051200-027E	Base frame, size 2, complete	1
3	HR51041100-027E	Back frame, size 1, complete	1
	HR51051100-027E	Back frame, size 2, complete	1
no illustr.	HR51045100E	Net upholstery, size 1, complete (incl. straps)	1
no illustr.	HR51055100E	Net upholstery, size 2, complete (incl. straps)	1
no illustr.	HR51045300	Trunk belt	1
no illustr.	HR51045140	Leg straps	2
no illustr.	HR51045200-103	Headrest, complete (accessory)	1

Hersteller/Manufacturer:

Otto Bock HealthCare GmbH

Max-Näder-Straße 15 · D-37115 Duderstadt

National: Telefon 05527-848-1461/1462/1463 · Fax 05527-848-14 60

International: Phone +49-5527-848-1304/1562/1590/1594/3663 · Fax +49-5527-848-1676

e-mail: reha@ottobock.de · Internet: <http://www.ottobock.com>

Versandanschrift für Rücksendungen/Address for Returns

Otto Bock HealthCare GmbH

Sälzerstraße 16 · D-56235 Ransbach-Baumbach



Otto Bock HealthCare GmbH has been certified by the German Society for the Certification of Quality Assurance Systems (DQS)
in accordance with DIN EN ISO 9001 standard, reg. no. 779 (management system)